

Keysight Technologies N5531S Measuring Receiver

Data Sheet



Unlocking Measurement Insights

Conditions and Requirements

The Keysight N5531S measuring receiver system is comprised of a PSA spectrum analyzer with Option 233, a P-Series or EPM Series power meter, and an N5532B/A¹ sensor module for metrology and calibration applications. To achieve the optimal measurement results as specified, the best metrology practice must be applied and the required instrument conditions must be met.

PSA is the core component instrument of the N5531S measuring receiver. The PSA instrument conditions included in the PSA specification guide must be met to meet the N5531S specifications.

Additional conditions required to meet specifications

- The system components are within their calibration cycle
- Tuned RF Level measurement is set to **High Accuracy Mode**
- Fast Mode is set to **Off** when performing modulation measurements
- For center frequency < 20 MHz, DC coupling is applied
- At least 2 hours of storage or operation at the operating temperature of 20 to 30 °C
- The PSA has been turned on at least 30 minutes with Auto Align On selected or if Auto Align Off is selected, Align All Now must be run:
 - Within the last 24 hours, and
 - Any time the ambient temperature changes more than 3 °C
 - After the analyzer has been at operating temperature at least 2 hours
- For analog modulation measurements, a direct connection between the PSA and the device under test (DUT) is required to achieve the best performance and meet the specifications for all test frequencies
- The following PSA options are required in addition to Option 233 as stated in the specifications:
 - Option 123 (pre-selector bypass) must be installed to meet TRFL specifications above 3 GHz
 - Option 107 (audio input 100 kΩ) is required with Option 233 (built-in measuring receiver personality) for the audio analysis
 - Option 1DS (pre-amplifier between 100 kHz and 3.05 GHz) or Option 110 (pre-amplifier between 10 MHz up to 50 GHz) is needed to achieve better sensitivity as indicated in the data sheet

Key features:

- Metrology-grade measurement accuracy
- Off-the-shelf, general-purpose instruments with specialized PSA measurement personality
- Best for signal source and step attenuator calibrations
- Abundant features with easy-to-use user interfaces
- Sensor modules covering up to 50 GHz with single input connection

Key measurements include:

- Frequency counter
- Absolute RF power
- Tuned RF level
- TRFL with tracking
- AM depth
- FM deviation
- oIM deviation
- Modulation rate
- Modulation distortion
- Modulation SINAD
- Audio frequency
- Audio AC level
- Audio distortion
- Audio SINAD
- Auto carrier triggering
- CCITT filters

1. As of April 1, 2010, the N5532A sensor module was discontinued and rolled to the N5532B. This document lists the N5532B/A as the N5532B and the N5532A share exactly the same specifications.

Key Specifications

For detailed specifications, see the “Measuring receiver personality” chapter in the PSA specifications guide (E4440-90647).

1.1 Frequency modulation

| Description | | Specification | Supplemental information | |
|--|------------------|---|---|--|
| Input power range ¹ | | -18 to +30 dBm | | |
| Operating rate range | | $100 \text{ kHz} \leq f_c < 10 \text{ MHz}$ | 20 Hz to 10 kHz | |
| | | $10 \text{ MHz} \leq f_c < 50 \text{ GHz}$ | 50 Hz to 200 kHz | |
| Peak frequency deviations ¹ | | | Peak deviation = IFBW/2 - Modulation rate | |
| | | $100 \text{ kHz} \leq f_c < 10 \text{ MHz}$ | 40 kHz maximum | |
| | | $10 \text{ MHz} \leq f_c < 50 \text{ GHz}$ | 400 kHz maximum | |
| | | | IFBW _{max} = 5 MHz in Auto mode; | |
| | | | IFBW _{max} = 10 MHz in Manual mode | |
| FM deviation accuracy ² | | | | |
| Frequency range | Modulation rate | Peak deviation | β^3 | |
| 250 kHz to 10 MHz | 20 Hz to 10 kHz | 200 Hz to 40 kHz | > 0.2 > 1.2 | $\pm 1.5\%$ of reading $\pm 1\%$ of reading |
| 10 MHz to 6.6 GHz | 50 Hz to 200 kHz | 250 Hz to 400 kHz | > 0.2 > 0.45 | $\pm 1.5\%$ of reading $\pm 1\%$ of reading |
| 6.6 to 13.2 GHz | 50 Hz to 200 kHz | 250 Hz to 400 kHz | > 0.2 > 8 | $\pm 2.5\%$ of reading $\pm 1\%$ of reading |
| 13.2 to 31.15 GHz | 50 Hz to 200 kHz | 250 Hz to 400 kHz | > 0.2 > 16 | $\pm 3.8\%$ of reading $\pm 1\%$ of reading |
| 31.15 to 50 GHz | 50 Hz to 200 kHz | 250 Hz to 400 kHz | > 0.2 > 32 | $\pm 8.5\%$ of reading $\pm 1\%$ of reading |
| AM rejection (50 Hz to 3 kHz BW) | | | | |
| Frequency range | Modulation rate | AM depths | | |
| 150 kHz to 3 GHz | 400 Hz or 1 kHz | $\leq 50\%$ | < 10 Hz peak deviation | |
| 3 to 6.6 GHz | 400 Hz or 1 kHz | $\leq 50\%$ | < 10 Hz | |
| 6.6 to 13.2 GHz | 400 Hz or 1 kHz | $\leq 50\%$ | < 20 Hz | |
| 13.2 to 26.5 GHz | 400 Hz or 1 kHz | $\leq 50\%$ | < 40 Hz | |
| 26.5 to 50 GHz | 400 Hz or 1 kHz | $\leq 50\%$ | < 75 Hz | |
| Residual FM (50 Hz to 3 kHz BW) | | | | |
| RF frequency | | | | |
| 100 kHz to 6.6 GHz | | < 1.5 Hz (rms) | | |
| 6.6 to 13.2 GHz | | < 3 Hz (rms) | | |
| 13.2 to 31.15 GHz | | < 6 Hz (rms) | | |
| 31.15 to 50 GHz | | < 12 Hz (rms) | | |
| Detectors | | | Available: +peak, -peak, \pm peak/2, peak hold, rms | |

1. The modulation rates and the peak deviations that the system is capable of measuring are governed by the instrument's IFBW

2. (Information Bandwidth) setting. Their relationship is described by the equation: Peak deviation (in Hz) = IFBW/2 - modulation rate.

3. When the carrier frequency f_c is less than 10 MHz, to avoid the 0 Hz frequency wrap-around, the f_c and IFBW must be chosen to satisfy $[f_c - (\text{IFBW})/2] > 100$ kHz.

4. β is the ratio of frequency deviation to modulation rate (deviation/rate).

Key Specifications (continued)

1.2 Amplitude modulation

| Description | | Specification | Supplemental information |
|---|------------------|---|---|
| Input power range | | -18 to +30 dBm | |
| Operating rate range ¹ | | $100 \text{ kHz} \leq f_c < 10 \text{ MHz}$ | 20 Hz to 10 kHz |
| | | $10 \text{ MHz} \leq f_c < 50 \text{ GHz}$ | 50 Hz to 200 kHz |
| Depth range | | 5 to 99 % | Capable of measuring AM depth range of 0 to 99 %. |
| AM depth accuracy² | | | |
| Frequency range | Modulation rate | Depths | |
| 100 kHz to 10 MHz | 50 Hz to 10 kHz | 5 % to 99 % | ±0.75 % of reading |
| 10 MHz to 3 GHz | 50 Hz to 100 kHz | 20 % to 99 % 5 % to 20 % | ±0.5 % of reading ±2.5 % of reading |
| 3 to 26.5 GHz | 50 Hz to 100 kHz | 20 % to 99 % 5 % to 20 % | ±1.5 % of reading ±4.5 % of reading |
| 26.5 to 31.15 GHz | 50 Hz to 100 kHz | 20 % to 99 % 5 % to 20 % | ±1.9 % of reading ±6.8 % of reading |
| 31.15 to 50 GHz | 50 Hz to 100 kHz | 20 % to 99 % 5 % to 20 % | ±6 % of reading ±26 % of reading |
| Flatness³ | | | |
| Frequency range | Modulation rate | Depths | |
| 10 MHz to 3 GHz | 90 Hz to 10 kHz | 5 % to 99 % | ±0.30 % of reading |
| 3 to 26.5 GHz | 90 Hz to 10 kHz | 5 % to 99 % | ±0.40 % of reading |
| 26.5 to 50 GHz | 90 Hz to 10 kHz | 5 % to 99 % | ±0.60 % of reading |
| FM rejection (50 Hz to 3 kHz BW) | | | |
| Frequency range | Modulation rate | Peak FM deviations | |
| 250 kHz to 10 MHz | 400 Hz or 1 kHz | < 5 kHz | < 0.14 % AM depth |
| 10 MHz to 50.0 GHz | 400 Hz or 1 kHz | < 50 kHz | < 0.36 % AM depth |
| Residual AM (50 Hz to 3 kHz BW) | | < 0.01 % (rms) ^{4,5} | |
| Detectors | | | Available: +peak, -peak, ±peak/2, peak hold, rms |

- When the carrier frequency f_c is less than 10 MHz, to avoid the 0 Hz frequency wrap-around, the f_c and IFBW must be chosen to satisfy $[f_c - (\text{IFBW})/2] > 100 \text{ kHz}$.
- For peak measurement only: AM accuracy may be affected by distortion generated by the measuring receiver. In the worst case this distortion can decrease accuracy by 0.1 % of reading for each 0.1 % of distortion.
- Flatness is the relative variation in indicated AM depth versus rate for a constant carrier frequency and depth.
- Preamp must be on to meet this specification for frequency range of 26.5 to 50 GHz.
- Follow this procedure to verify this specification: Input a clean CW signal (0 dBm) to the measuring receiver; Manually tune the frequency to the input signal; Set the PSA parameters as follows, (1) IF BW = 6 kHz, (2) Detector type = RMS, (3) High Pass Filter = 50 Hz, (4) Low Pass Filter = 3 kHz, (5) Set **RF Input Ranging** to **Man**, and decrease the input attenuation at 2 dB/step until **SigHi** message appears, and then back off 2 dB for the **SigHi** message to disappear.

Key Specifications (continued)

1.3 Phase modulation

| Description | | Specification | Supplemental information |
|--|---|-----------------------------|--------------------------|
| Input power range | | -18 to +30 dBm | |
| Operating rate range | $100 \text{ kHz} \leq f_c < 10 \text{ GHz}$ | 200 Hz to 20 kHz | |
| Maximum peak phase deviation | $f_c < 10 \text{ MHz}$ | 450 radians ¹ | |
| | $f_c \geq 10 \text{ MHz}$ | 12,499 radians ² | In Auto mode |
| | | 24,999 radians ² | In Manual mode |
| Φ M accuracy | | | |
| Frequency range | Deviations | | |
| 100 kHz to 6.6 GHz | > 0.7 rad | ± 1 % of reading | |
| | > 0.3 rad | ± 3 % of reading | |
| 6.6 to 13.2 GHz | > 2.0 rad | ± 1 % of reading | |
| | > 0.6 rad | ± 3 % of reading | |
| 13.2 to 26.5 GHz | > 4.0 rad | ± 1 % of reading | |
| | > 1.2 rad | ± 3 % of reading | |
| 26.5 to 31.5 GHz | > 4.0 rad | ± 1 % of reading | |
| | > 1.3 rad | ± 3 % of reading | |
| 31.5 to 50 GHz | > 8.0 rad | ± 1 % of reading | |
| | > 2.4 rad | ± 3 % of reading | |
| AM rejection (50 Hz to 3 kHz BW) | For 50 % AM at 1 kHz rate | < 0.03 rad (peak) | |
| Residual PM (50 Hz to 3 kHz BW) | | | |
| Frequency range | | | |
| 100 kHz to 6.6 GHz | < 0.0017 rad (rms) | | |
| 6.6 to 13.2 GHz | < 0.0033 rad (rms) | | |
| 13.2 to 31.15 GHz | < 0.0066 rad (rms) | | |
| 31.15 to 50 GHz | < 0.0130 rad (rms) | | |
| Detectors | Available: +peak, -peak, \pm peak/2, peak hold, rms | | |

- When the carrier frequency f_c is less than 10 MHz, to avoid the 0 Hz frequency wrap-around, the f_c and IFBW must be chosen to satisfy $[f_c - (\text{IFBW}/2)] > 100 \text{ kHz}$. The specification of 450 radians applies for $f_c = 200 \text{ kHz}$, IFBW = 200 kHz, and a modulation rate of 200 Hz. The specification for maximum peak phase deviation will linearly improve as the allowed IFBW increase. As f_c increases, the IFBW can increase up to the maximum allowed IFBW in **Auto** or **Manual** modes.
- When the carrier frequency (f_c) is equal to or greater than 10 MHz, the maximum peak deviation that the instrument is capable of measuring depends on the IFBW setting and the modulation rate of the signal-under-test. The relationship is described by the equation:

$$\text{Max peak deviation (in radians)} = [\text{IFBW}/(2 \times \text{modulation rate in Hz})] - 1$$
- The maximum IFBW used in **Auto** mode is $5 \times 10^6 \text{ Hz}$, therefore, Max peak deviation (in radians) = $(2.5 \times 10^6 / \text{modulation rate in Hz}) - 1$. In **Manual** mode, the maximum IFBW can be set to 107 Hz, hence, Max peak deviation (in radians) = $(5 \times 10^6 / \text{modulation rate in Hz}) - 1$.

Key Specifications (continued)

1.4 RF frequency counter

| Description | Specification | Supplemental information |
|-----------------------------------|---------------------------------|---|
| Range | 100 kHz to 50 GHz | In Auto mode |
| Sensitivity ¹ | | |
| 100 kHz $\leq f_c < 3.0$ GHz | 0.4 mV _{rms} (-55 dBm) | |
| 3.0 GHz $\leq f_c \leq 26.5$ GHz | 1.3 mV _{rms} (-45 dBm) | |
| 26.5 GHz $\leq f_c \leq 50$ GHz | 4.0 mV _{rms} (-35 dBm) | |
| Maximum resolution | 0.001 Hz | |
| Accuracy | | |
| Modes | | Frequency and frequency error (manual tuning) |
| Sensitivity in manual tuning mode | | Using manual ranging and changing RBW settings, sensitivity can be increased to approximately -100 dBm. |

1.5 Audio input²

| Description | Specification | Supplemental information |
|--------------------------|------------------------------|--------------------------|
| Frequency range | 20 Hz to 250 kHz | |
| Input impedance | | 100 k Ω (nominal) |
| Maximum safe input level | 7 V _{rms} or 20 VDC | |

1.6 Audio frequency counter

| Description | Specification | Supplemental information |
|-----------------------|---|--|
| Frequency range | 20 Hz to 250 kHz | |
| Accuracy ³ | | With HPF set to minimum setting of < 20 Hz |
| $f < 1$ kHz | $\pm(0.02 \text{ Hz} + f \times \text{Internal Reference Accuracy})^4$ | |
| $f \geq 1$ kHz | ± 3 counts of the first 6 significant digits $\pm f \times (\text{Internal Reference Accuracy})$ | |
| Resolution | 0.01 Hz (8 digits) | |
| Sensitivity | ≤ 5 mV | |

1.7 Audio AC (RMS) level

| Description | Specification | Supplemental information |
|-------------------------|---|--------------------------|
| Frequency range | 20 Hz to 250 kHz | |
| Measurement level range | 100 mV _{rms} to 3 V _{rms} | |
| Accuracy | 1 % of reading | |
| Detector mode | | RMS |

- Instrument condition: RBW ≤ 1 kHz.
- All audio measurements require PSA Option 107.
- Follow this procedure to verify this specification: Set an input audio signal at 100 mV. Set the PSA as follows: (1) Auto Level, (2) Auto IF BW, (3) LP is greater than the audio frequency, (4) HP = 300 Hz or less than the audio frequency, and (5) Average = 5 Repeat.
- See the "Internal Time Base Reference" section in the PSA specification guide for the Internal Reference Accuracy.

1.8 Audio distortion

| Description | Specification | Supplemental information |
|-------------------------------------|-----------------------------|-------------------------------------|
| Display range (20 Hz to 250 kHz BW) | 0.01 to 100 % (-80 to 0 dB) | |
| Accuracy (20 Hz to 250 kHz) | ±1 dB of reading | |
| Residual noise and distortion | < 0.3 % (-50.4 dB) | |
| Total noise | | -73.2 dB characteristic performance |
| Total distortion | | -74.8 dB characteristic performance |

1.9 Audio SINAD

| Description | Specification | Supplemental information |
|-------------------------------------|-------------------|------------------------------------|
| Display range (20 Hz to 250 kHz BW) | 0.00 to 80 dB | |
| Display resolution | 0.01 dB | |
| Accuracy | | |
| 20 Hz to 20 kHz | ±1 dB of reading | |
| 20 kHz to 250 kHz | ±2 dB of reading | |
| Residual noise and distortion | 50.4 dB (< 0.3 %) | |
| Total noise | | 73.2 dB characteristic performance |
| Total distortion | | 74.8 dB characteristic performance |

1.10 Audio filters

| Description | Specification | Supplemental information |
|---|---|---|
| Filter flatness | | |
| Non high-pass filter | | < ± 1 % at rates > 20 Hz |
| 50 Hz high-pass filter | < ±1 % at rates > 50 Hz | |
| 300 Hz high-pass filter | < ±1 % at rates > 300 Hz | |
| 400 Hz high-pass filter ¹ | < ±1 % at rates > 400 Hz | |
| 3 kHz low-pass filter | < ±1 % at rates < 3,030 Hz | |
| 15 kHz low-pass filter | < ±1 % at rates < 15,030 Hz | |
| 30 kHz low-pass filter ¹ | < ±1 % at rates < 30 kHz | |
| 80 kHz low-pass filter ¹ | < ±1 % at rates < 80 kHz | |
| > 100 kHz low-pass filter | < ± 1 % at rates < 100 kHz | |
| CCITT weighting filter | CCITT recommendation P53 | |
| Deviation from the ideal | ±0.2 dB at 800 Hz; ±1.0 dB, 300 Hz to 3 kHz | |
| CCITT filter response | ±2.0 dB, 50 to 300 Hz and 3 to 3.5 kHz; ±3.0 dB, 3.5 to 5 kHz | |
| De-emphasis filters | 25 μs, 50 μs, 75 μs, and 750 μs | De-emphasis filters are single-pole, low-pass filters with nominal -3 dB frequencies of: 6,366 Hz for 25 μs, 3,183 Hz for 50 μs, 2,122 Hz for 75 μs, and 212 Hz for 750 μs. |
| Deviation from ideal de-emphasis filter | < 0.4 dB, or < 3° | Applicable to 25 μs, 50 μs, and 75 μs filters. With 3 kHz Low-Pass filter and IFBW Mode set to minimal. |

1. PSA firmware revision ≥ A.11.08 or Option 23B

1.11 RF Power^{1,2}

The Keysight N5531S measuring receiver system with the N5532B/A sensor modules performs RF power measurements from -10 dBm (100 μ W) to +30 dBm (1 W). The N5531S must be used with Keysight P-Series power meters (N1911A, N1912A), or EPM Series power meters (N1913A, N1914A). A LAN/GPIB gateway will be required if the EPM/EPM-P Series power meter is used.

| Description | Specification | | | | Supplemental information | | | | |
|--------------------------|--------------------------------|---------------|-------------|-------------|--------------------------|-------------|-------------|-------------|-------------|
| RF power accuracy (dB) | | | | | | | | | |
| | Power meter range 1 | | | | Typicals | | | | |
| | Sensor module options | | | | Sensor module options | | | | |
| | +20 to +30 dBm | 504 | 518 | 526 | 550 | 504 | 518 | 526 | 550 |
| | 100 kHz $\leq f_c \leq$ 10 MHz | ± 0.356 | – | – | – | ± 0.182 | – | – | – |
| | 10 MHz $< f_c \leq$ 30 MHz | ± 0.356 | ± 0.361 | – | – | ± 0.182 | ± 0.185 | – | – |
| | 30 MHz $< f_c \leq$ 2 GHz | ± 0.356 | ± 0.361 | ± 0.361 | ± 0.361 | ± 0.182 | ± 0.185 | ± 0.185 | ± 0.185 |
| | 2 GHz $< f_c \leq$ 4.2 GHz | ± 0.356 | ± 0.392 | ± 0.422 | ± 0.387 | ± 0.182 | ± 0.201 | ± 0.217 | ± 0.188 |
| | 4.2 GHz $< f_c \leq$ 18 GHz | – | ± 0.400 | ± 0.422 | ± 0.387 | – | ± 0.205 | ± 0.217 | ± 0.188 |
| | 18 GHz $< f_c \leq$ 26.5 GHz | – | – | ± 0.480 | ± 0.387 | – | – | ± 0.217 | ± 0.199 |
| | 26.5 GHz $< f_c \leq$ 50 GHz | – | – | – | ± 0.420 | – | – | – | ± 0.216 |
| | Power meter range 2 | | | | Typicals | | | | |
| | Sensor module options | | | | Sensor module options | | | | |
| | 0 to +20 dBm | 504 | 518 | 526 | 550 | 504 | 518 | 526 | 550 |
| | 100 kHz $\leq f_c \leq$ 10 MHz | ± 0.190 | – | – | – | ± 0.097 | – | – | – |
| | 10 MHz $< f_c \leq$ 30 MHz | ± 0.190 | ± 0.200 | – | – | ± 0.097 | ± 0.101 | – | – |
| | 30 MHz $< f_c \leq$ 2 GHz | ± 0.190 | ± 0.200 | ± 0.200 | ± 0.200 | ± 0.097 | ± 0.101 | ± 0.101 | ± 0.101 |
| | 2 GHz $< f_c \leq$ 4.2 GHz | ± 0.190 | ± 0.255 | ± 0.301 | ± 0.212 | ± 0.097 | ± 0.130 | ± 0.154 | ± 0.108 |
| | 4.2 GHz $< f_c \leq$ 18 GHz | – | ± 0.267 | ± 0.301 | ± 0.212 | – | ± 0.136 | ± 0.154 | ± 0.108 |
| | 18 GHz $< f_c \leq$ 26.5 GHz | – | – | ± 0.380 | ± 0.247 | – | – | ± 0.195 | ± 0.126 |
| | 26.5 GHz $< f_c \leq$ 50 GHz | – | – | – | ± 0.297 | – | – | – | ± 0.152 |
| RF power resolution | Display resolution | 0.001 dB | | | | | | | |
| Instrumentation accuracy | Logarithmic | ± 0.02 dB | | | | | | | |
| | Linear | ± 0.5 % | | | | | | | |

1. For latest specification updates refer to N1911A/N1912A and N1913A/N1914A power meter User's Guides.

2. The N5531S RF Power Accuracy is derived from the power meter accuracy. The parameters listed in this section are components used to calculate the RF Power Accuracy. Fundamentals of RF and Microwave Power Measurements (5988-9215EN) does an excellent job of explaining how the components are combined to derive an overall accuracy number. The resulting calculation yields ± 0.190 to ± 0.297 dB when measuring a +10 dBm signal and ignoring DUT mismatch. Assuming 1.5:1 DUT SWR, the calculation would return a typical accuracy of ± 0.213 to ± 0.387 dB (depending on the frequency range and power under test). Absolute and relative accuracy specifications do not include mismatch uncertainty.

1.11 RF Power (continued)

| Description | Specification | Supplemental information |
|---|---------------------------------------|---------------------------------------|
| Input SWR | | |
| N5532B/A Option 504 | 100 kHz to 2 GHz | < 1.10:1 ($\rho = 0.048$) |
| | 2 GHz to 4.2 GHz | < 1.28:1 ($\rho = 0.123$) |
| N5532B/A Option 518 | 10 MHz to 2 GHz | < 1.10:1 ($\rho = 0.048$) |
| | 2 GHz to 18 GHz | < 1.28:1 ($\rho = 0.123$) |
| N5532B/A Option 526 | 30 MHz to 2 GHz | < 1.10:1 ($\rho = 0.048$) |
| | 2 GHz to 18 GHz | < 1.28:1 ($\rho = 0.123$) |
| | 18 GHz to 26.5 GHz | < 1.40:1 ($\rho = 0.167$) |
| N5532B/A Option 550 | 30 MHz to 2 GHz | < 1.10:1 ($\rho = 0.048$) |
| | 2 GHz to 18 GHz | < 1.28:1 ($\rho = 0.123$) |
| | 18 GHz to 26.5 GHz | < 1.40:1 ($\rho = 0.167$) |
| | 26.5 GHz to 33 GHz | < 1.55:1 ($\rho = 0.216$) |
| | 33 GHz to 40 GHz | < 1.70:1 ($\rho = 0.259$) |
| | 40 GHz to 50 GHz | < 1.75:1 ($\rho = 0.272$) |
| Zero set and measurement noise ¹ | N5532B/A ± 680 nW | |
| Zero drift of sensors (1 hour, at constant temperature after 24 hour warm-up) | N5532B/A < ± 100 nW | |
| RF power ranges of N5531S with N5532B/A sensor modules | -20 dBm (10 μ W) to +30 dBm (1 W) | One range for power sensors |
| Response time (0 to 99 % of reading) | | 150 ms x number of averages (nominal) |
| Displayed units | Watts, dBm, or Volts | |

1. Since zero set and measurement noise cannot be separated, these two components are combined as one error term.

Key Specifications (continued)

TRFL Specification Nomenclature

The tuned RF level measurement uncertainty is represented primarily by two regions. For high signal-to-noise (S/N) measurements, the uncertainty is dominated by the linearity of the measuring receiver. For low S/N measurements, the measurement uncertainty is dominated by the noise of the measuring receiver being added to the measured signal. The input power at which the uncertainty switches from linearity dominated to noise dominated is labeled as Residual noise threshold. The minimum power level is defined as the noise floor of the measuring receiver system.

Additionally, there are 2 range-to-range change uncertainties known as Range 2 Uncertainty and Range 3 Uncertainty, respectively. Range 2 Uncertainty occurs when the measuring receiver switches from Range 1 to Range 2, and Range 3 uncertainty from Range 2 to Range 3. They are additive uncertainties applied to all measurements whose input powers across Range Switch Level.




Figure 1. Measurement uncertainty vs. input power relationship

Key Specifications (continued)

1.12 Tuned RF level^{1, 2, 3}

Note: While the Tuned RF level specifications listed below are for IFBW settings of 75 Hz and 10 Hz, the IFBW in N5531S can also be set to 30 kHz or 200 kHz. The wider IFBW is capable of measuring sources with some degree of frequency instability by trading off measurement sensitivity.

For sources with frequency instability greater than 100 kHz, use the Tuned RF Level with Tracking measurement. When using the Tuned RF Level with Tracking, the following additional amplitude error must be applied due to FFT frequency response as the signal drifts within the tracking range: $\pm(0.15 \text{ dB} + 0.1 \text{ dB/MHz of span})$ to a max of $\pm 0.40 \text{ dB}$, where span is equivalent to the tracking range setting in the measurement. The Tuned RF Level with Tracking measurement upper frequency limit = 3.05 GHz. For the Tuned RF Level with Tracking, the minimum power in the = $10 * \log [\text{Integrated BW}/(75 \text{ Hz} * 1.06)]$, relative to the specified 75 Hz minimum power level.

| Description | Specification | | Supplemental information | | | |
|---|------------------------|---------------------------|-------------------------------------|----------------------------------|-------------------------------------|-----------------------------------|
| Power range | | | | | | |
| Maximum power | Preamp off | +30 dBm | | | | |
| | Preamp on | +16 dBm | | | | |
| Minimum power (dBm)⁷ E4443A/45A/40A | Frequency range | 75 Hz IFBW | | 10 Hz IFBW^{4, 5} | | Also see Notes 1 and 2 on page 15 |
| | | Preamp uninstalled | Preamp installed⁶ | Preamp uninstalled | Preamp installed⁶ | |
| | 100 kHz to 2 MHz | -110 | -124/-110 | -129 | -140/-129 | |
| | 2 to 10 MHz | -115 | -131/-115 | -134 | -140/-134 | |
| | 10 MHz to 3.05 GHz | -117 | -134/-133 | -136 | -140/-140 | |
| | 3.05 to 6.6 GHz | -117 | -117/-127 | -136 | -136/-140 | |
| | 6.6 to 13.2 GHz | -108 | -108/-116 | -127 | -127/-135 | |
| | 13.2 to 19.2 GHz | -100 | -100/-110 | -119 | -119/-129 | |
| 19.2 to 26.5 GHz | -93 | -93/-102 | -112 | -112/-121 | | |
| Minimum power (dBm)⁷ E4447A/46A/48A | Frequency range | 75 Hz IFBW | | 10 Hz IFBW^{4, 5} | | Also see Notes 1 and 2 on page 15 |
| | | Preamp uninstalled | Preamp installed⁶ | Preamp uninstalled | Preamp installed⁶ | |
| | 100 kHz to 2 MHz | -110 | -124/-110 | -129 | -140/-129 | |
| | 2 to 10 MHz | -115 | -131/-115 | -134 | -140/-134 | |
| | 10 MHz to 3.05 GHz | -117 | -134/-133 | -136 | -140/-140 | |
| | 3.05 to 6.6 GHz | -114 | -114/-126 | -133 | -133/-140 | |
| | 6.6 to 13.2 GHz | -111 | -111/-123 | -130 | -130/-140 | |
| | 13.2 to 19.2 GHz | -109 | -109/-118 | -128 | -128/-137 | |

1. PSA Option 123 is required to perform Tuned RF Level measurements above 3 GHz.
2. These specifications are valid when the measuring receiver input is a CW tone and operating temperature is within the range of 20 to 30 °C.
3. Absolute and relative accuracy specifications do not include mismatch uncertainty.
4. With 10 Hz IFBW setting selected, the measurement automatically switches the RBW to the 1 Hz setting for SNR values < 10 dB.
5. For an instrument with a serial number prefix below US/MY4615, the minimum power level in 10 Hz IFBW setting is 10 dB higher than the values shown here. However, if the PSA contains Option 107, the values shown in the table still apply.
6. In the frequency range of 100 kHz to 3.05 GHz, the minimum power specifications with Preamp installed are presented in two values: A/B, where value A is for the PSA installed with Option 1DS, and value B is for the PSA installed with Option 110. Furthermore, in the frequency range of 100 kHz and 10 MHz, Option 110 is turned off for these measurements. Option 1DS only covers a frequency range of 100 kHz and 3.05 GHz, whereas Option 110 covers up to the maximum frequency of the PSA base instrument. Those two preamplifier options cannot coexist in the same PSA instrument.
7. With 30 kHz and 200 kHz IF bandwidth (IFBW), TRFL minimum power level will be degraded by a factor of $10 \log(\text{IFBW}/75 \text{ Hz})$, relative to the specified 75 Hz minimum power level. This will result in a degradation of 26 dB for the 30 kHz IFBW and 34 dB for the 200 kHz IFBW.

Key Specifications (continued)

1.12 Tuned RF level (continued)

| Description | | Specification | | | | Supplemental information |
|---|--|---|----------------------------------|-----------------------|---------------------|-----------------------------------|
| Minimum power (dBm) E4447A/46A/48A | | 75 Hz IFBW | | 10 Hz IFBW | | Also see Notes 1 and 2 on page 14 |
| | Frequency range | Preamp uninstalled | Preamp installed ⁶ | Preamp uninstalled | Preamp installed | |
| | 19.2 to 26.5 GHz | -97 | -97/-104 | -116 | -116/-123 | |
| | 26.5 to 31.15 GHz | -98 | -98/-103 | -117 | -117/-122 | |
| | 31.15 to 41 GHz | -87 | -87/-91 | -106 | -106/-110 | |
| | 41 to 45 GHz | -81 | -81/-81 | -100 | -100/-100 | |
| | 45 to 50 GHz | -69 | -69/-69 | -88 | -88/-88 | |
| Linearity | | $\pm(0.009 \text{ dB} + 0.005 \text{ dB}/10 \text{ dB step})^1$ | | | | |
| Relative measurement accuracy | Residual noise threshold ² to maximum power | $\pm(0.015 \text{ dB} + 0.005 \text{ dB}/10 \text{ dB step})^{1,3,4}$ (nominal) | | | | |
| | Minimum power to residual noise threshold | $\pm(\text{cumulative error}^e + 0.0012 \times (\text{input power} - \text{residual noise threshold power})^2)$ | | | | |
| Residual noise threshold power (dBm) | | Residual noise threshold power = minimum power +30 dB | | | | |
| Range 2 uncertainty ⁶ | | $\pm 0.031 \text{ dB}$ | | | | |
| Range 3 uncertainty ⁷ | | $\pm 0.031 \text{ dB}$ | | | | |

- Step in this specification refers to the difference between relative measurements, such as might be experienced by stepping a stepped attenuator. Therefore, accuracy is computed by adding the uncertainty for each full or partial 10 dB step to the other uncertainty term. For example, if the two levels whose relative level is to be determined differ by 15 dB, consider that to be a difference of two 10-dB steps.
- The residual noise threshold power is the power level at which the signal-to-noise ratio (SNR) becomes the dominant contributor to the measurement uncertainty. See TRFL Specifications Nomenclature at the beginning of this section.
- Immediately following the system alignments, the measurement is made by manually setting frequency to that of the signal-under-test, Accuracy mode to High, and Measure Control to Single. For the E4446A/E4447A/E4448A, if the change of measured frequency crosses frequency bands (refer to previous page in the column of "Supplemental Information" for definitions of frequency bands for the E4446A/E4447A/E4448A), allow 10 minutes for thermal stability before taking the first measurement within the new band.
- This includes the linearity accuracy.
- In relative accuracy of TRFL measurements, the cumulative error is the error incurred when stepping from a higher power level to the Residual Noise Threshold Power level. The formula to calculate the cumulative error is $\pm(0.015 \text{ dB} + 0.005 \text{ dB}/10 \text{ dB step})$. For example, assume the higher level starting power is 0 dBm and the calculated Residual Noise Threshold Power is -99 dBm. The cumulative error would be $\pm(0.015 + 99/10 \times 0.005 \text{ dB})$, or $\pm 0.065 \text{ dB}$, where x is a ceiling function that means the smallest integer is not less than x.
- Add this specification when the measuring receiver enters the Range 2 state. Range 2 is entered when the Range 1 signal-to-noise ratio (SNR) falls between 50 and 28 dB. The SNR value is tuning band dependent. A prompt of Range 2 in the PSA display will indicate that the measuring receiver is in Range 2.
- Add this specification in addition to Range 2 Uncertainty when the measuring receiver software enters the Range 3 state. Range 3 is entered when the Range 2 SNR falls between 50 and 28 dB. The SNR value is tuning band dependent. A prompt of Range 3 in the PSA display will indicate that the measuring receiver is in Range 3.

Key Specifications (continued)

1.12 Tuned RF level (continued)

| Description | Specification | Supplemental information |
|--------------------------------------|---|---|
| Absolute measurement accuracy | Preamp Off | |
| | +20 dBm to maximum power | $\pm(\text{power meter range 1 uncertainty} + 0.005 \text{ dB}/10 \text{ dB step})$ |
| | Residual noise threshold power to +20 dBm | $\pm(\text{power meter range 2} - 4 \text{ uncertainty} + 0.005 \text{ dB}/10 \text{ dB step})$ |
| | Minimum power to residual noise threshold power | $\pm(\text{cumulative error} + 0.0012 \times (\text{input power} - \text{residual noise threshold power})^2)$ |
| Absolute measurement accuracy | Preamp On | |
| | Residual noise threshold power to +16 dBm | $\pm(\text{power meter range 2} - 4 \text{ uncertainty} + 0.005 \text{ dB}/10 \text{ dB step})$ |
| | Minimum power to residual noise threshold power | $\pm(\text{cumulative error}^1 + 0.0012 \times (\text{input power} - \text{residual noise threshold power})^2)$ |
| Residual Noise Threshold Power (dBm) | | Residual Noise Threshold Power = Minimum Power + 30 dB |
| Range 2 Uncertainty ² | $\pm 0.031 \text{ dB}$ | |
| Range 3 Uncertainty ³ | $\pm 0.031 \text{ dB}$ | |

- In absolute accuracy of TRFL measurements, the cumulative error is the error incurred when stepping from a higher power level to the Residual Noise Threshold power level. See Figure 1 for a graphic. In order to calculate the cumulative error, you must determine the Residual Noise Threshold power and determine the Power Meter Range. The formula to calculate the cumulative error is: $\pm(\text{Power Meter Range Uncertainty} + 0.005 \text{ dB}/10 \text{ dB step})$. For example: the power sensor is Option 504, starting power is 0 dBm and power will be stepped to -120 dBm. Therefore starting power falls in the Power Meter Range 2-4 and the uncertainty is $\pm 0.190 \text{ dB}$, as indicated in the table on the next page. The Residual Noise Threshold Power is -106 dBm at IFBW of 10 Hz. This is calculated per the Minimum Power specification in the table on the previous page. Assume no preamp is installed, and that the measurement frequency is 10 MHz to 3 GHz. The Residual Noise Threshold Power is $-136 \text{ dBm} + 30 \text{ dB} = -106 \text{ dBm}$ using the formula on this page. The cumulative error is then $\pm(0.190 \text{ dB} + 106/10 \times 0.005 \text{ dB})$, or $\pm 0.245 \text{ dB}$, where x is a ceiling function that means the smallest integer not less than x, which is 11 in this example.
- Add this specification when the Measuring Receiver enters the Range 2 state. Range 2 is entered when the Range 1 signal-to-noise ratio (SNR) falls between 50 and 28 dB. The SNR value is tuning band dependent. A prompt of Range 2 in the PSA display will indicate that the Measuring Receiver is in Range 2.
- Add this specification in addition to Range 2 Uncertainty when the Measuring Receiver enters the Range 3 state. Range 3 is entered when the Range 2 SNR falls between 50 and 28 dB. The SNR value is tuning band dependent. A prompt of Range 3 in the PSA display will indicate that the Measuring Receiver is in Range 3.

NOTE

As the displayed average noise level (DANL) of a spectrum analyzer becomes very low, it can reveal residuals. These occur at discrete frequencies and arise from the various clocks and other components of the local oscillators. This is true for all modern spectrum analyzers. The residuals specification for the PSA Series is -100 dBm. Please take this information into consideration when you measure the TRFL level below -100 dBm. A user may apply a 50-ohm terminator to the PSA's RF input connector and switch to the PSA's spectrum analysis mode to verify the PSA residuals.

The sensor module (N5532B/A) may generate a residual of around -100 dBm or lower at frequency of 50 MHz and its harmonics. Please take this information into consideration when you use the N5532B/A to measure the TRFL level below -100 dBm at 50 MHz and its harmonics.

Key Specifications (continued)

1.12 Tuned RF level (continued)

| Description | Specification | | | | Supplemental information | | | |
|--------------------------------|---|-------------|-------------|-------------|--------------------------|-------------|-------------|-------------|
| Power Meter Range Uncertainty | | | | | | | | |
| | Power meter Uncertainty (dB) | | | | Typicals | | | |
| | Sensor module options | | | | Sensor module options | | | |
| +20 to +30 dBm | 504 | 518 | 526 | 550 | 504 | 518 | 526 | 550 |
| 100 kHz $\leq f_c \leq$ 10 MHz | ± 0.356 | — | — | — | ± 0.182 | — | — | — |
| 10 MHz $< f_c \leq$ 30 MHz | ± 0.356 | ± 0.361 | — | — | ± 0.182 | ± 0.185 | — | — |
| 30 MHz $< f_c \leq$ 2 GHz | ± 0.356 | ± 0.361 | ± 0.361 | ± 0.361 | ± 0.182 | ± 0.185 | ± 0.185 | ± 0.185 |
| 2 GHz $< f_c \leq$ 4.2 GHz | ± 0.356 | ± 0.392 | ± 0.422 | ± 0.367 | ± 0.182 | ± 0.201 | ± 0.217 | ± 0.188 |
| 4.2 GHz $< f_c \leq$ 18 GHz | — | ± 0.400 | ± 0.422 | ± 0.367 | — | ± 0.205 | ± 0.217 | ± 0.188 |
| 18 GHz $< f_c \leq$ 26.5 GHz | — | — | ± 0.480 | ± 0.367 | — | — | ± 0.217 | ± 0.199 |
| 26.5 GHz $< f_c \leq$ 50 GHz | — | — | — | ± 0.420 | — | — | — | ± 0.216 |
| | Power meter range 2¹ Uncertainty (dB) | | | | Typicals | | | |
| | Sensor module Options | | | | Sensor module Options | | | |
| 0 to +20 dBm | 504 | 518 | 526 | 550 | 504 | 518 | 526 | 550 |
| 100 kHz $\leq f_c \leq$ 10 MHz | ± 0.190 | — | — | — | ± 0.097 | — | — | — |
| 10 MHz $< f_c \leq$ 30 MHz | ± 0.190 | ± 0.200 | — | — | ± 0.097 | ± 0.101 | — | — |
| 30 MHz $< f_c \leq$ 2 GHz | ± 0.190 | ± 0.200 | ± 0.200 | ± 0.200 | ± 0.097 | ± 0.101 | ± 0.101 | ± 0.101 |
| 2 GHz $< f_c \leq$ 4.2 GHz | ± 0.190 | ± 0.255 | ± 0.301 | ± 0.212 | ± 0.097 | ± 0.130 | ± 0.154 | ± 0.108 |
| 4.2 GHz $< f_c \leq$ 18 GHz | — | ± 0.267 | ± 0.301 | ± 0.212 | — | ± 0.136 | ± 0.154 | ± 0.108 |
| 18 GHz $< f_c \leq$ 26.5 GHz | — | — | ± 0.380 | ± 0.247 | — | — | ± 0.195 | ± 0.126 |
| 26.5 GHz $< f_c \leq$ 50 GHz | — | — | — | ± 0.297 | — | — | — | ± 0.152 |
| Operating frequency range | | | | | | | | |
| E4443A/45A/40A/47A/46A/48A | 100 kHz to 3 GHz | | | | | | | |
| E4443A/45A/40A/47A/46A/48A | 3 to 6.7 GHz | | | | Requires Option 123 | | | |
| E4445A/40A/47A/46A/48A | 6.7 to 13.2 GHz | | | | Requires Option 123 | | | |
| E4440A/47A/46A/48A | 13.2 to 26.5 GHz | | | | Requires Option 123 | | | |
| E4447A/46A/48A | 26.5 to 42.98 GHz | | | | Requires Option 123 | | | |
| E4446A/48A | 42.98 to 44 GHz | | | | Requires Option 123 | | | |
| E4448A | 44 to 50 GHz | | | | Requires Option 123 | | | |
| Displayed units | | | | | | | | |
| | Absolute | | | | Watts, dBm, or Volts | | | |
| | Relative | | | | Percent or dB | | | |
| Displayed resolution | | | | | | | | |
| | 6 digits in Watts or 5 digits in Volts mode | | | | | | | |
| | 0.001 dB in dBm or dB (relative) mode | | | | | | | |
| Input SWR | | | | | | | | |
| | See "RF Power" Section | | | | | | | |

1. Refer to the PSA specification guide, E4440-90647, for more information.

N5531S Ordering Information

The Keysight N5531S measuring receiver system is comprised of a PSA, a P-Series or EPM Series power meter, and an N5532B sensor module.

PSA Series spectrum analyzer

(Select one model from the following models)

- E4443A
3 Hz to 6.7 GHz
- E4445A
3 Hz to 13.2 GHz
- E4440A
3 Hz to 26.5 GHz
- E4447A
3 Hz to 42.98 GHz
- E4446A
3 Hz to 44 GHz
- E4448A
3 Hz to 50 GHz

PSA options (x = 0, 3, 5, 6, 7, 8)

E444xA-233

Built-in measuring receiver personality and PC software (required)

E444xA-123

Switchable preselector bypass (required for TRFL measurements above 3 GHz)

E444xA-1DS

RF internal preamplifier (required for the best TRFL specifications up to 3.05 GHz; does not co-exist with Option 110)

E444xA-110

RF/ μ W internal preamplifier (required for the best TRFL specifications up to the maximum frequency of the PS base instrument; does not co-exist with Option 1DS)

E444xA-107

Audio input 100 k Ω (required for audio analysis, only operational with Option 233)

AM/FM/PM triggering

Shipped standard with Option 233 (PSA firmware rev \geq A.11.08) or Option 23A

CCITT filter (adds CCITT and 400-Hz HP, 30-kHz/80-kHz LP filters)

Shipped standard with Option 233 (PSA firmware rev \geq A.11.08) or Option 23B

Select from PSA options for other measurements (Optional, Refer to PSA Configuration Guide for details of option compatibility and requirements)

PSA option upgrades¹ (x = 0, 3, 5, 6, 7, 8)

E444xAU-233

Built-in measuring receiver personality and PC software (required)

E444xAU-123

Switchable preselector bypass (required for TRFL measurements above 3 GHz)

E444xAU-1DS

RF internal preamplifier (required for the best TRFL specifications up to 3.05 GHz; does not co-exist with Option 110)

E444xAU-110

RF/ μ W internal preamplifier (required for the best TRFL specifications up to the maximum frequency of the PSA base instrument; does not co-exist with Option 1DS)

E444xAU-107

Audio input 100 k Ω (required for audio analysis, only operational with Option 233)

(Select one from the following models)

P-Series power meter

N1911A

P-Series single channel power meter

N1912A

P-Series dual channel power meter

EPM Series power meter

N1913A

EPM Series single channel power meter

N1914A

EPM Series dual channel power meter

1. Upgrades for certain PSA options may not be available for earlier instruments. For detailed information regarding availability and compatibility of options, please visit http://www.keysight.com/find/psa_upgrades

Select from power meter options (optional)

N5532B sensor module

(Select one frequency option)

N5532B-504

100 kHz to 4.2 GHz, type-N (m) input connector

N5532B-518

10 MHz to 18 GHz, type-N (m) input connector

N5532B-526

30 MHz to 26.5 GHz, APC-3.5 (m) input connector

N5532B-550

30 MHz to 50 GHz, 2.4 mm (m) input connector

N5532B-019

Adaptor to N191xA power meter (required when the N191xA power meter is used), can also be ordered standalone

Select from N5532B options (optional)

Accessories

N5531S-010

LAN connection kit (including one LAN hub and 3 regular LAN cables) (optional)

Note: For existing N5531S measuring receiver users who buy the N5532B sensor module to replace the N5532A: Using the N5532B in place of the N5532A requires the following firmware versions:

- E444xA PSA: \geq A.11.21
- N1911A power meter: \geq A.05.02
- N1912A power meter: \geq A.05.02
- N1913A power meter: \geq A.01.06
- N1914A power meter: \geq A.01.06


Related Literature

| Publication title | Publication type | Publication number |
|--|-------------------------|--------------------|
| N5531S measuring receiver | | |
| N5531S Measuring Receiver | Technical Overview | 5989-4795EN |
| Accurate Absolute and Relative Power Measurement Using the Keysight N5531S Measuring Receiver System | Application Note | 5989-8161EN |
| PSA in general | | |
| Spectrum Analyzer and Signal Analyzer | Selection Guide | 5968-3413E |
| Power meter in general | | |
| P-Series Power Meters and Power Sensors | Data Sheet | 5989-2471EN |
| P-Series Power Meters and Power Sensors | Technical Overview | 5989-1049EN |
| EPM Series Power Meters | Data Sheet | 5990-4019EN |
| Power measurement fundamentals | | |
| Fundamentals of RF and Microwave Power Measurements, Introduction to Power, History, Definition, International Standards, and Traceability | Application Note 1449-1 | 5988-9213EN |
| Fundamentals of RF and Microwave Power Measurements, Power Sensors and Instrumentation | Application Note 1449-2 | 5988-9214EN |
| Fundamentals of RF and Microwave Power Measurements, Power Measurement Uncertainty per International Guides | Application Note 1449-3 | 5988-9215EN |
| Fundamentals of RF and Microwave Power Measurements, An Overview of Agilent Instrumentation for RF/Microwave Power Measurement | Application Note 1449-4 | 5988-9216EN |

Evolving Since 1939

Our unique combination of hardware, software, services, and people can help you reach your next breakthrough. We are unlocking the future of technology.

From Hewlett-Packard to Agilent to Keysight.



For more information on Keysight Technologies' products, applications or services, please contact your local Keysight office. The complete list is available at: www.keysight.com/find/contactus

Americas

| | |
|---------------|------------------|
| Canada | (877) 894 4414 |
| Brazil | 55 11 3351 7010 |
| Mexico | 001 800 254 2440 |
| United States | (800) 829 4444 |

Asia Pacific

| | |
|--------------------|----------------|
| Australia | 1 800 629 485 |
| China | 800 810 0189 |
| Hong Kong | 800 938 693 |
| India | 1 800 11 2626 |
| Japan | 0120 (421) 345 |
| Korea | 080 769 0800 |
| Malaysia | 1 800 888 848 |
| Singapore | 1 800 375 8100 |
| Taiwan | 0800 047 866 |
| Other AP Countries | (65) 6375 8100 |

Europe & Middle East

| | |
|----------------|---------------|
| Austria | 0800 001122 |
| Belgium | 0800 58580 |
| Finland | 0800 523252 |
| France | 0805 980333 |
| Germany | 0800 6270999 |
| Ireland | 1800 832700 |
| Israel | 1 809 343051 |
| Italy | 800 599100 |
| Luxembourg | +32 800 58580 |
| Netherlands | 0800 0233200 |
| Russia | 8800 5009286 |
| Spain | 800 000154 |
| Sweden | 0200 882255 |
| Switzerland | 0800 805353 |
| | Opt. 1 (DE) |
| | Opt. 2 (FR) |
| | Opt. 3 (IT) |
| United Kingdom | 0800 0260637 |

For other unlisted countries: www.keysight.com/find/contactus (BP-9-7-17)

DEKRA Certified
ISO 9001 Quality Management System

www.keysight.com/go/quality
Keysight Technologies, Inc.
DEKRA Certified ISO 9001:2015
Quality Management System

myKeysight

myKeysight

www.keysight.com/find/mykeysight

A personalized view into the information most relevant to you.

http://www.keysight.com/find/emt_product_registration

Register your products to get up-to-date product information and find warranty information.

KEYSIGHT SERVICES

Accelerate Technology Adoption.
Lower costs.

Keysight Services

www.keysight.com/find/service

Keysight Services can help from acquisition to renewal across your instrument's lifecycle. Our comprehensive service offerings—one-stop calibration, repair, asset management, technology refresh, consulting, training and more—helps you improve product quality and lower costs.



Keysight Assurance Plans

www.keysight.com/find/AssurancePlans

Up to ten years of protection and no budgetary surprises to ensure your instruments are operating to specification, so you can rely on accurate measurements.

Keysight Channel Partners

www.keysight.com/find/channelpartners

Get the best of both worlds: Keysight's measurement expertise and product breadth, combined with channel partner convenience.



Unlocking Measurement Insights

This information is subject to change without notice.
© Keysight Technologies, 2017
Published in USA, December 1, 2017
5989-9217EN
www.keysight.com